



Replacement Oxygen Sensors for Respiratory and Anaesthesia applications

The Maxtec® comprehensive selection of oxygen sensors fulfills virtually any customer application. Our standard (alkaline-based) sensors meet or exceed OEM specifications for respiratory applications. The Max-250 series uses a patented weak-acid based technology to provide a more stable signal, longer sensor life, and withstand high levels of CO₂, CO, and other acidic gases.



The Maxtec range of Sensors

MAX-1 (12 month warranty).....R113P90 Replacement for GE (Datex-Ohmeda) (common form factor).	MAX-14 (14 month warranty).....R116P80 Replacement for MSA MiniOX 1A.
MAX-2 (18 month warranty).....R107P20 Extra-Life Replacement for Drager Narkomed Series. Large sensor replaces O ₂ assembly cap and sensor.	MAX-16 (24 month warranty).....R114P70 Replacement for Covidien/PB Ventilators
MAX-3 (12 month warranty).....R109P11 Replacement for Paragon Platinum SC430 and Penion Prima SP2.	MAX-16HT (12 month warranty).....R114P73 Replacement for Versamed Vent.
MAX-7 (12 month warranty).....R109P09 Replacement for Teledyne AX/MX300 Series, R17MED/#C-43690.	MAX-17 (12 month warranty).....R116P10 Replacement for Teledyne Monitor 60T, 191T and 2000T.
MAX-8 (18 month warranty).....R107P22 Extra-Life Replacement for GE(Datex-Ohmeda) . For external applications only. Large sensor replaces O ₂ assembly cap and sensor.	MAX-18 (12 month warranty).....R116P40 Replacement for Teleflex Medical (Hudson RCI) 5568.
MAX-9 (14 . warranty).....R114P80 Replacement for Teleflex Medical (Hudson RCI) 5500. (CE available upon request).	MAX-19 (12 month warranty).....R116P60 Replacement for Teleflex Medical (HudsonRCI) 5569.
MAX-10 (12 month warranty).....R112P10 Threaded replacement for GE (Datex-Ohmeda) 7900 Smart Vent Series Ventilator.	MAX-23 (12 month warranty).....R116P06 Replacement for Teledyne R22, #C-44611.
MAX-11 (12 month warranty).....R113P10 Dual Cathode replacement for Drager Narkomed Series (common form factor) oxygen analyser with internal sensor.	MAX-43 (12 month warranty).....R109P07 Replacement for GE (Datex-Ohmeda) Giraffe.
MAX-11i (12 month warranty).....R113P06 Dual Cathode replacement for Drager incubators, IC8000 (High temperature applications).	MAX-47 (12 month warranty).....R112P16 Replacement for Teleflex Medical (Hudson RCI) 5800 Monitor, 5801 Analyser.
MAX-12 (14 month warranty).....R109P45 Replacement for Siemens 300 and 900 Series.	MAX-250+ (24 month warranty).....R125P02-011 Oxygen replacement sensor for the MAXO ₂ +A
MAX-13 (14 month warranty).....R115P10 Replacement for MSA MiniOX 1, 11, 111, 3000 PB O ₂ Monitor.	MAX-250 (24 month warranty).....R125P01-002 Oxygen replacement sensor for the OM-25A.
MAX-13-250 (24 month warranty).....R125P07 Extra-Life Replacement for MSA MiniOX 1, 11, 111, 3000, PB O ₂ Monitor.	MAX-250E (24 month warranty).....R125P03-002 Oxygen replacement sensor for the MAXO ₂ +AE, OM-25ME, OM-25AE, MAXBlend, MaxVenturi and Centry Blender.
	MAX-250ESF (12 month warranty).....R115P85 Oxygen replacement sensor. Small size and fast response. MAXO ₂ +AE,OM-25AE and MAXBlend.

Technology House, Chowley Oak, Tattenhall, Chester, Cheshire. CH3 9GA

Tel: 01829 771792

Fax: 01829 771791

Email: www.envinsci.co.uk